

1.0 Reference and Address			
Report Number	181202323SHA-001	Original Issued: 18-Feb-2019	Revised: None
Standard(s)	Medical Electrical Equipment - Part 1: General Requirements For Basic Safety And Essential Performance [AAMI ES60601-1:2005 +A1] Medical Electrical Equipment - Part 1: General Requirements For Basic Safety And Essential Performance [CSA C22.2#60601-1:2014 Ed.3] Medical Electrical Equipment - Part 1: General Requirements For Basic Safety & Essential Performance [IEC 60601-1:2005 Ed.3+A1]		
Applicant	Applied power (Shanghai) Electronics Co., Ltd	Manufacturer	Dongguan YingJu Power Co.,LTD
Address	No.400 Fangchun Rd,free trade zone Shanghai,China 201203	Address	No. 6 YongXing RD, Shayao Village, Shijie Town , Dongguan City,Guangdong,523300,China.
Country	China	Country	China
Contact	John Song	Contact	Jin Chun Hai
Phone	13524367623	Phone	0769-81837999
FAX	--	FAX	--
Email	jsong@appliedpsu.com	Email	jinchunhai@yingju.cn

2.0 Product Description	
Product	Medical Power Supply
Brand name	APPLIED POWER
Description	The product covered by this report is medical power supply with various input and output types. It is intended to be used below 5,000 m altitude, indoor use, Class II. Operation Ambient Temperature: -10°C to 40°C Operation Humidity: 20% to 95%RH non-condensing The insulation was evaluated at 2MOPP.
Models	MIA-11 followed by U,C,A,or K; followed by A,B or C.
Model Similarity	The first follow can be "U" or "C", "A", "K", which donates different input type ("U" for USA and Japan, "C" for China, "A" for Australia, "K" for United kingdom) The second follow can be "A", "B" or "C", which donates different output type ("A" for USB type A, "B" for Output Cable with TBD barrel connector, and "C" for USB type C) The models covered in this report only differ from the rating output and the non-critical components thereof.
Ratings	Input:100-240~,50-60Hz,0.4A Output:5Vdc,2.1A,10.5W
Other Ratings	NA