



Ref. Certif. No.

**SE-100943**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product

Medical Power Supply

Name and address of the applicant

Applied Power (Shanghai) Electronics Co., Ltd.  
No.400 Fangchun Rd. free trade zone, Shanghai, 201203,  
China

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Note: When more than one factory, please report on page 2

Chenzhou Frecom Electronics Co., Ltd.  
1/F, No.7 Bldg. Chenzhou Hengfeng Science and Technology Co.  
Ltd. Hanning West Road, Jiangbei Industrial Park, Zixing  
Economic Development Zone, Chenzhou City, 431081, China

Ratings and principal characteristics

Input: 100-240V~, 50-60Hz, 1.0A  
Output: 5V@20W, 9V@27W, 12V-48V@30W, Class II

Trademark (if any)

APPLIED POWER

Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

MIS30B\*\*\*\*

Additional information (if necessary may also be reported on page 2)

See page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1

As shown in the Test Report Ref. No. which forms part of this Certificate

200300695SHA-001

This CB Test Certificate is issued by the National Certification Body

**Intertek Semko AB**  
Torshamnsgatan 43  
Box 1103  
SE-164 22 Kista, Sweden

**intertek**

Signature:

Gary Hu

Date: 30 April, 2020



Ref. Certif. No.

**SE-100943**

**Additional information (if necessary)**

The national differences for Canada, USA have been checked.

Explanation of MIS30B\*\*\*\*:

The first symbol "\*\*\*" can be 05, 09, 12, 15, 18, 24, 36, 48, represents different output voltage.

The second symbol "\*\*\*" can be AA, AB, AC, AE, AF, BA, BB, BC, BE, BF, CA, CB, CC, CE, CF, represents internal control code.

Date: 30 April, 2020

Signature: 